

Spec Clean™ Ultrasonic Cleaner Solution

Description:

Spec Clean ultrasonic cleaner solution is a super concentrated solution formulated for use in ultrasonic cleaning devices. The product is available in gallons with metered pump. It may be diluted with ordinary tap water.

Enhances Ultrasonic Cleaning Action:

- Penetrating agents improve cleaning action in small crevices and hinges.
- Odor suppressant neutralizes organic and chemical odors.

Protects Delicate Instruments:

- Team of anti-corrosive agents protects metal surfaces.
- Best results for carbon steel items including burs and curets.
- Safe cleaning for jointed orthodontic pliers.
- Enhances instrument appearance and shine.

Meet CDC Guidelines for Instrument Decontamination:

- Effective cleaning is essential prior to sterilization.
- Effective ultrasonic cleaning reduces risk of injury during manual scrubbing of instruments.

Spec Clean ultrasonic cleaner solution is a biodegradable instrument cleaner that meets the demand for more environmentally friendly products.



Clean and protect against corrosion.



Combines great cleaning action with excellent anti-corrosives and instrument brighteners.

Product Specifications

pH Concentrate	8.5-10.5
Use Temperature	68°F (20°C) to 140°F (60°C). Do not exceed 140°F (60°C). Avoid running continuous cycles that may overheat the cleaning solution
Odor	Unscented
Shelf Life	2 years from date of manufacture
Disposal	Dispose according to all local, state and federal regulations. This product contains no phosphates

Description	Case Qty	Order#
1 Gallon with Pump	4	GVUS128-1

Intended Use:

Spec Clean is intended for cleaning instruments by aiding the cavitation process of ultrasonic cleaners. It is designed to minimize manual handling/scrubbing of contaminated items and is especially useful for cleaning instruments with serrated edges, joints and hinges. It will safely clean delicate instruments, anodized aluminum and carbon steel curets or burs

Product Capabilities:

Spec Clean rapidly removes organic and inorganic soils. The detergent formula lifts and disperses soils during the ultrasonic cleaning process.

A powerful anti-corrosive system protects delicate instruments including carbon steel and anodized aluminum. The anti-corrosive agents are especially effective to protect box locks in jointed items and hinges.

Use Precautions:

Do not add other chemicals to Spec Clean. Change the cleaning solution daily or more frequently if soiled or cloudy. Empty the ultrasonic tank, rinse and wipe out completely DAILY. Do not allow residue to accumulate. Follow ultrasonic equipment manufacturer's recommendations for cleaning and maintenance.

Due to the corrosive effects of water, it is not recommended to soak metal items for prolonged periods (over two hours or overnight).

Spec Clean is an eye irritant and mild skin irritant. Follow CDC Guidelines for protective attire during instrument decontamination: safety goggles and/or face shield, mask (for biological hazards) and water-resistant gloves. Store in cool, dry area away from direct sunlight. Solution will fade when exposed to direct sunlight.

Availability:

Stocked by MicroCare Medical distributors. Contact MicroCare Medical for a distributor near you or visit MicroCare.com

MicroCare Medical
6120 East 58th Avenue
Commerce City, CO 80022
(800) 843-3343



Visit Us:
MicroCare.com